



EMSL Analytical, Inc.

2205 Corporate Plaza Parkway SE, Suite 200 Smyrna, GA 30080

Phone/Fax: (770) 956-9150 / (770) 956-9181

<http://www.EMSL.com> / atlantalab@emsl.com

EMSL Order: 071905494
Customer ID: BYPG34
Customer PO:
Project ID:

Attention: Lawrence Transve
Buyers Protection Group
4300 Alexander Drive
Suite 200
Alpharetta, GA 30022
Project: Doe / 123456

Phone: (610) 417-0763
Fax:
Collected Date: 07/08/2019
Received Date: 07/09/2019 8:00 AM
Analysis Date: 07/09/2019

Test Report: Total Coliform E.coli (P/A) (Presence/Absence)

Method: SM-9223 B 20th Ed. for Water Samples

Sample	Sample Location Date/Time Collected	Amount Received mL	Amount Sampled mL	Total Coliform Present/Absent	E. coli Present/Absent
071905494-0001 1	Kitchen 07/08/2019 12:00PM	100	100	Absent	Absent

Analyst(s): _____

Michael Murphy (1)

Michael Murphy
or other approved signatory

Colilert® can detect coliforms at 1 CFU/100 mL after 24 hours of incubation. Colilert®-18 can detect coliforms at 1 CFU/100 mL analyzed after 18 hours of incubation; NA = not applicable. High levels (>2 million) of non-coliform bacteria may produce false-positive results. EMSL maintains liability limited to cost of analysis. Interpretation and use of test results are the responsibility of the client. This report relates only to the samples reported above, and may not be reproduced, except in full, without written approval by EMSL. EMSL bears no responsibility for sample collection activities or analytical method limitations. The report reflects the samples as received. When the information supplied by the customer can affect the validity of the result, it will be noted on the report.
Samples analyzed by EMSL Analytical, Inc Smyrna, GA

Initial report from: 07/16/2019 09:03:52